**Performance**

Measuring range	20 to 500 $\mu\text{g}/\text{m}^3$	500 to 1200 $\mu\text{g}/\text{m}^3$
Sampling Rate	100 ml/min (3000 ml)	100 ml/min (1500 ml)
Correction factor	1	2.4
Sampling time	30 min	15 min

Detecting limit : 5 $\mu\text{g}/\text{m}^3$ (3000ml)



Colour change : Yellow \rightarrow Purple

Corrections for temperature : Necessary for 5 to 40°C

Corrections for humidity : Unnecessary for R.H. 20 to 80 %.

Relative standard deviation : 10 % (for 20 to 100 $\mu\text{g}/\text{m}^3$), 5 % (for 100 to 500 $\mu\text{g}/\text{m}^3$)

Shelf life : 2 years

Reaction principle**Possible coexisting substances and their interferences** (NOTE : Page 2-5)

Substance	Interference	Changes colour by itself to
Hydrogen chloride, Chlorine	No	No
Vinyl chloride	+	Purple
1,2-Dichloroethylene	+	Purple
Tetrachloroethylene	+	Purple
1,1,1-Trichloroethane	No	No
Toluene	No	No
Xylene	No	No

Calibration gas generation

Permeation tube method

Special note

In case of outdoor measurement, keep the tube out of direct sunlight.