

# GASTEC Instructions for No.105 Higher Class Hydrocarbons Detector Tube

## FOR SAFE OPERATION :

Read this manual and the instruction manual of your Gastec Gas Sampling Pump carefully.

### ⚠ WARNING:

1. Use only Gastec detector tubes in a Gastec Pump.
2. Do not interchange or use non-Gastec parts or components in Gastec's detector tube and pump system.
3. The use of non-Gastec parts or components in Gastec's detector tube and pump system or use of a non-Gastec detector tube with a Gastec pump or use of a Gastec detector tube with a non-Gastec pump may result in property damage, serious bodily injury, and death; voids all warranties; and voids all performance and data accuracy guaranties.

### ⚠ CAUTION : If not observed, injuries to the operator or damage to the product may result.


1. When breaking the tube ends, keep away from eyes.
2. Do not touch the broken glass tubes, pieces and reagent with bare hand(s).
3. The sampling time represents the time necessary to draw the air sample through the tube. The tube must be positioned in the desired sampling area for the entire sampling time or until the flow finish indicator indicates the end of the sample.

### △ NOTES : For maintaining performance and reliability of the test result

1. Use Gastec Gas Sampling Pump together with Gastec Detector Tubes only for the purposes specified in the instruction manual of the detector tube.
2. Use this tube under the temperature range of 0 - 40°C (32 - 104°F).
3. Use this tube under the relative humidity range of 0 - 90%.
4. This tube may be interfered by the coexisting gases. Please refer to the "INTERFERENCES".
5. Shelf life and storage conditions of the tube is marked on the label of the box of tube.

**APPLICATION OF THE TUBE :** Use of this tube for the detection of Higher Class Hydrocarbons in air or the industrial areas and environmental atmospheric condition.

**SPECIFICATION :** (As a result of Gastec's commitment to continued improvement, specifications are subject to change without notice.)



Measuring Range	100 - 200 ppm	200 - 3,000 ppm
Number of Pump Strokes	2	1
Correction Factor	1/2	1
Sampling Time	1.5 minutes per pump stroke	
Detecting Limit	20 ppm (n=2)	
Color Change	White → Blackish brown	
Reaction Principle	Higher class hydrocarbons reduce iodine pentoxide to liberate iodine to produce blackish brown in color.	

**Coefficient of Variation 10% (for 200 to 1,000ppm), 5% (for 1,000 to 3,000 ppm)**

**\*\* Shelf Life : Please refer to the Validity Date printed on the box of tube.**

**\*\* Store the tubes in dark and cool place.**

## CORRECTION FOR TEMPERATURE, HUMIDITY AND PRESSURE :

**Temperature :** Temperature correction is not required.

**Humidity :** Humidity correction is not required.

**Pressure :** To correct for pressure, multiply the tube reading by

$$\frac{\text{Tube Reading (ppm)} \times 1013 \text{ (hPa)}}{\text{Atmospheric Pressure (hPa)}}$$

## MEASUREMENT PROCEDURE :

1. For leak checking of the pump insert a fresh sealed detector tube into pump. Follow instructions provided with the pump operating manual.
2. Break tips off a fresh detector tubes and a scrubber tube in the tube tip breaker of the pump.
3. Insert the scrubber tube into the pump inlet with arrow (➔) on the tube pointing toward pump.
4. Make certain pump handle is all the way in. Align guide marks on pump body and handle.
5. Pull the handle all the way out until it locks on 1 pump stroke (100ml). Wait 1.5 minutes and confirm the completion of the sampling.
6. For lower than 200 ppm measurement, repeat the above sampling procedure one more time.
7. Read concentration at interface of the stained-to-unstained reagent.
8. If atmospheric correction is needed, refer to the "Corrections for Pressure, and Pump strokes.

## INTERFERENCES :

Substance	Concentration	Interference	Change color by itself
Acetylene, Ethylene	0.1% or higher	Plus error	Produce blackish brown
Carbon monoxide	0.1% or higher	Plus error	Produce blackish brown
Other organic solvents		Plus error	Produce blackish brown

## APPLICATION FOR OTHER GASES :

Tube 105 can also be used for other substance as below:

Substance	Correction Factor	No. of Pump Strokes	Measuring Range
Decane	2.0	1 or 2	200 - 6,000 ppm
Heptane	0.9	1 or 2	90 - 2,700 ppm
Hexane	0.8	1 or 2	80 - 2,400 ppm
Nonane	1.3	1 or 2	130 - 3,900 ppm
Octane	1.0	1 or 2	100 - 3,000 ppm

**CORRECTION FACTOR :** Detector tubes are primarily designed to measure specific gases. But it is also possible to measure other substances of similar chemical properties with the aid of a correction factor or chart. Therefore, please make use of the correction factor/chart measuring ranges as a reference. For a more precise factor please contact your Gastec distributor.

## DISPOSAL INSTRUCTION :

Reagent of the tubes use toxic chromic acid. On disposing the tube regardless of used or unused, follow the rules and regulations of the local government.

**WARRANTY :** If you have any questions regarding gas detection and quality of the tube, please feel free to contact your Gastec representatives.