

GASTEC Instructions for No.1HH Carbon Monoxide Extra High Range Detector Tube

FOR SAFE OPERATION :

Read this manual and the instruction manual of your Gastec Gas Sampling Pump carefully.

⚠ WARNING:

1. Use only Gastec detector tubes in a Gastec Pump.
2. Do not interchange or use non-Gastec parts or components in Gastec's detector tube and pump system.
3. The use of non-Gastec parts or components in Gastec's detector tube and pump system or use of a non-Gastec detector tube with a Gastec pump or use of a Gastec detector tube with a non-Gastec pump may result in property damage, serious bodily injury, and death; voids all warranties; and voids all performance and data accuracy guaranties

⚠ CAUTION : If not observed, injuries to the operator or damage to the product may result.


1. When breaking the tube ends, keep away from eyes.
2. Do not touch the broken glass tubes, pieces and reagent with bare hand(s).
3. The sampling time represents the time necessary to draw the air sample through the tube.
The tube must be positioned in the desired sampling area for the entire sampling time or until the flow finish indicator indicates the end of the sample.

△ NOTES : For maintaining performance and reliability of the test result

1. Use Gastec Gas Sampling Pump together with Gastec Detector Tubes only for the purposes specified in the instruction manual of the detector tube.
2. Use this tube within the temperature range of 0 - 40°C (32 - 104°F).
3. Use this tube within the relative humidity range of 0 - 90%.
4. This tube may be interfered by the coexisting gases. Please refer to the "INTERFERENCES".
5. Shelf life and storage conditions of the tube are marked on the label of the box of tube.

APPLICATION OF THE TUBE : Use of this tube for the detection of Carbon Monoxide in air or the industrial areas and environmental atmospheric condition.

SPECIFICATION : (As a result of Gastec's commitment to continued improvement, specifications are subject to change without notice.)



Measuring Range	1 - 2 %	2 - 50 %
Number of Pump Strokes	1	1/2
Correction Factor	1/2	1
Sampling Time	1 minute per 1/2 pump stroke	
Detecting Limit	0.5 % (n = 1)	
Color Change	White → Blackish Brown	
Reaction Principle	Carbon monoxide reacts with iodine pentoxide to produce blackish brown stain.	

Coefficient of Variation : 10% (for 2 to 15 %), 5 % (for 15 to 50%)

**** Shelf Life : Please refer to the Validity Date printed on the box of tube.**

**** Store the tubes in dark and cool place.**

CORRECTION FOR TEMPERATURE, HUMIDITY AND PRESSURE :

Calibration of the Gastec detector Tube No. 1 HH is based on a tube temperature of 20° (68° F) and not the temperature of the gas being sampled, approximately 50 % relative humidity and normal atmospheric pressure.

Temperature : Temperature correction is not required for 0 - 40°C (32 - 104°F)

Humidity : Humidity correction is not required for 0 - 90 %.

Pressure : To correct for pressure, multiply the tube reading by

$$\frac{\text{Tube Reading (\%)} \times 1013 \text{ (hPa)}}{\text{Atmospheric Pressure (hPa)}}$$

MEASUREMENT PROCEDURE :

1. For leak checking of the pump insert a fresh sealed detector tube into pump.
Follow instructions provided with the pump operating manual.
2. Break tips off a fresh detector tube in the tube tip breaker of the pump.
3. Insert the tube into the pump inlet with arrow **G** on the tube pointing toward pump.
4. Make certain pump handle is all the way in. Align guide marks on pump body and handle.
5. Pull the handle up to half pump stroke(50 ml). Wait 1 minute and confirm the completion of the sampling.
6. For lower than 2 % measurement, prepare fresh tube then pull full pump stroke. Obtain true concentration by multiplying the correction factor of 1/2.
7. Read concentration at the interface of the stained-to-unstained reagent.
8. If correction is needed, multiply the correction factor of pump stroke and Pressure.

INTERFERENCES :

Substance	Concentration	Interference	Change color by itself
Acetylene	≥ 0.2 %	Plus error	Produce blackish brown
Olefins(Ethylene, Propylene)	≥ 2/3	~	Produce brown stain.
Butane	≥ 1 %	~	Produce pale brown for whole layer
Propane	≥ 20 %	~	

The table of this interference gases primarily expresses the interference of each coexisting gas in the gas concentration range, equivalent to the gas concentration. Therefore, the test result may be given positive result by the other substance not listed in the table. For more precise information is needed, please contact us or our distributors in your territory.

DANGEROUS AND HAZARDOUS PROPERTIES :

Threshold Limit Value-Time Weighted Average by ACGIH (2002) : 25 ppm

DISPOSAL INSTRUCTION :

Reagent of the tube use chromium and selenium. On disposing the tube regardless of whether used or unused, follow the rules and regulations of the local government.

WARRANTY :

If you have any questions regarding gas detection and quality of the tubes, please feel free to contact your Gastec representatives.

Manufacturer : Gastec Corporation
6431 Fukaya, Ayase-City, 252-1103, Japan

IM001HHE1
Printed in Japan
02K1Z